### §821.4

- (i) The lot or batch within which a device was manufactured;
- (ii) The serial number of a specific device:
- (iii) The expiration date of a specific device;
- (iv) The date a specific device was manufactured.
- (v) For an HCT/P regulated as a device, the distinct identification code required by \$1271.290(c) of this chapter.

[58 FR 43447, Aug. 16, 1993, as amended at 67 FR 5951, Feb. 8, 2002; 78 FR 58822, Sept. 24, 2013]

#### §821.4 Imported devices.

For purposes of this part, the importer of a tracked device shall be considered the manufacturer and shall be required to comply with all requirements of this part applicable to manufacturers. Importers must keep all information required under this part in the United States.

## Subpart B—Tracking Requirements

## §821.20 Devices subject to tracking.

- (a) A manufacturer of any class II or class III device that fits within one of the three criteria within §821.1(a) must track that device in accordance with this part, if FDA issues a tracking order to that manufacturer.
- (b) When responding to premarket notification submissions and remarket approval applications, FDA will notify the sponsor by issuing an order that states that FDA believes the device meets the criteria of section 519(e)(1) of the act and, by virtue of the order, the sponsor must track the device.

[67 FR 5951, Feb. 8, 2002]

# § 821.25 Device tracking system and content requirements: manufacturer requirements.

- (a) A manufacturer of a tracked device shall adopt a method of tracking for each such type of device that it distributes that enables a manufacturer to provide FDA with the following information in writing for each tracked device distributed:
- (1) Except as required by order under section 518(e) of the act, within 3 working days of a request from FDA, prior to the distribution of a tracked device

- to a patient, the name, address, and telephone number of the distributor, multiple distributor, or final distributor holding the device for distribution and the location of the device:
- (2) Within 10 working days of a request from FDA for tracked devices that are intended for use by a single patient over the life of the device, after distribution to or implantation in a patient:
- (i) The unique device identifier (UDI), lot number, batch number, model number, or serial number of the device or other identifier necessary to provide for effective tracking of the devices;
- (ii) The date the device was shipped by the manufacturer;
- (iii) The name, address, telephone number, and social security number (if available) of the patient receiving the device, unless not released by the patient under §821.55(a);
- (iv) The date the device was provided to the patient;
- (v) The name, mailing address, and telephone number of the prescribing physician;
- (vi) The name, mailing address, and telephone number of the physician regularly following the patient if different than the prescribing physician; and
- (vii) If applicable, the date the device was explanted and the name, mailing address, and telephone number of the explanting physician; the date of the patient's death; or the date the device was returned to the manufacturer, permanently retired from use, or otherwise permanently disposed of.
- (3) Except as required by order under section 518(e) of the act, within 10 working days of a request from FDA for tracked devices that are intended for use by more than one patient, after the distribution of the device to the multiple distributor:
- (i) The unique device identifier (UDI), lot number, batch number, model number, or serial number of the device or other identifier necessary to provide for effective tracking of the devices;
- (ii) The date the device was shipped by the manufacturer;
- (iii) The name, address, and telephone number of the multiple distributor: